



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

VM-C 2023 Viral Markers

INSTITUTION: LKF Laboratorium fur Klinische Forschung GmbH
Schwentental GE 24223

ATTENTION: Volker El-Samalouti PhD

CAP NUMBER: 7234136-01 **Kit#** 1

KIT INFORMATION:	Kit ID:	Kit Mailed:	Original Evaluation:	Next Mailing Date:
	36439454	8/21/2023	9/25/2023	1/16/2024

COPIED TO: CAP

LEGEND: Exception Reason Codes appearing in this evaluation:
<NONE>

S. [Signature]

Reviewed By

27/11/2023

Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 7234136-01

Kit# 1

Institution: LKF Laboratorium fur Klinische Forschung GmbH

Attention: Volker El-Samalouti PhD

City / State: Schwentimental GE 24223

Kit ID: 36439454

Kit Mailed: 8/21/2023

Original Evaluation: 9/25/2023

**EVALUATION
ORIGINAL**

VM-C 2023 Viral Markers

Test Method	Specimen	Your Result	Intended Response	Your Grade
HBsAg ROCHE c600 SER/E170	VM1-27	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-28	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable
	VM1-29	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-30	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-31	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
Anti-HBs ROCHE c600 SER/E170	VM1-27	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-28	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-29	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable
	VM1-30	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-31	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 7234136-01

Kit# 1

Institution: LKF Laboratorium für Klinische Forschung GmbH

Attention: Volker El-Samalouti PhD

City / State: Schwentental GE 24223

Kit ID: 36439454

Kit Mailed: 8/21/2023

Original Evaluation: 9/25/2023

**EVALUATION
ORIGINAL**

VM-C 2023 Viral Markers

Test Method	Specimen	Your Result	Intended Response	Your Grade
Anti-HBc Total (IgM an ROCHE c600 SER/E170	VM1-27	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-28	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-29	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable
	VM1-30	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-31	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable
Anti-HAV Total (IgM an ROCHE c600 SER/E170	VM1-27	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable
	VM1-28	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable
	VM1-29	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable
	VM1-30	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-31	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 7234136-01

Kit# 1

Institution: LKF Laboratorium fur Klinische Forschung GmbH

Attention: Volker El-Samalouti PhD

City / State: Schwentimental GE 24223

Kit ID: 36439454

Kit Mailed: 8/21/2023

Original Evaluation: 9/25/2023


EVALUATION

ORIGINAL

VM-C 2023 Viral Markers

Test Method	Specimen	Your Result	Intended Response	Your Grade
Anti-HIV 1/2 Assay ROCHE e600 SER/E170	VM1-27	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable
	VM1-28	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-29	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-30	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-31	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable
Anti-HCV Screen ROCHE e600 SER/E170	VM1-27	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable
	VM1-28	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable
	VM1-29	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable
	VM1-30	NON-REACTIVE/NEGATIVE	NON-REACTIVE/NEGATIVE	Acceptable
	VM1-31	REACTIVE/POSITIVE	REACTIVE/POSITIVE	Acceptable

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 7234136-01

Kit# 1

Institution: LKF Laboratorium für Klinische Forschung GmbH

Attention: Volker El-Samalouti PhD

City / State: Schwentimental GE 24223

Kit ID: 36439454

Kit Mailed: 8/21/2023

Original Evaluation: 9/25/2023

**EVALUATION
ORIGINAL**

VM-C 2023 Viral Markers

CAP

CAP #: 7234136

Subspecialty: General Immunology

Regulated Analyte	Proficiency Event 2023 1			Proficiency Event 2023 2			Proficiency Event 2023 3			Current Event Performance Interpretation	Cumulative CLIA '88 Performance Interpretation
	Test Event	Score	%	Test Event	Score	%	Test Event	Score	%		
Alpha-1 Antitrypsin										Pending	
Alpha-Fetoprotein										Pending	
Anti-HIV	VM-A	5/5	100	VM-B	5/5	100	VM-C	5/5	100	Satisfactory	Successful
Complement C3										Pending	
Complement C4										Pending	
Hepatitis (Anti-HBc)	VM-A	5/5	100	VM-B	5/5	100	VM-C	5/5	100	Satisfactory	Successful
Hepatitis (HBsAg)	VM-A	5/5	100	VM-B	5/5	100	VM-C	5/5	100	Satisfactory	Successful
IgA										Pending	
IgE	K-A	5/5	100							Pending	Pending <4>
IgG										Pending	
IgM										Pending	
General Immunology		20/20	100		15/15	100		15/15	100	Satisfactory	Successful

<4> Scorecard performance pending future evaluation or may not be applicable due to discontinued testing or use of a waived method.

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.