



COLLEGE of AMERICAN
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

LS-B 2023 Ligand Special

INSTITUTION: LKF Laboratorium fur Klinische Forschung GmbH
Schwentinental GE 24223


ATTENTION: Volker El-Samalouti PhD

CAP NUMBER: 7234136-01 **Kit#** 1

KIT INFORMATION:	Kit ID:	Kit Mailed:	Original Evaluation:	Next Mailing Date:
	36415123	9/18/2023	10/26/2023	4/22/2024

COPIED TO: CAP

LEGEND: Exception Reason Codes appearing in this evaluation:
<NONE>



Reviewed By



Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 7234136-01

Kit# 1

Institution: LKF Laboratorium fur Klinische Forschung GmbH

Attention: Volker El-Samalouti PhD

City/ State: Schwentimental GE 24223

Kit ID: 36415123

Kit Mailed: 9/18/2023

Original Evaluation: 10/26/2023

**EVALUATION
ORIGINAL**

LS-B 2023 Ligand Special

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics									Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+ 100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper		Your Grade	
Insulin uIU/mL ROCHE e600 SER/E170	ING-04	229.1	223.68	11.80	86	+0.5	188.2	259.1	Acceptable	
	ING-05	18.5	18.45	0.98	86	0.0	15.5	21.4	Acceptable	
	ING-06	64.8	64.09	2.86	85	+0.3	55.5	72.7	Acceptable	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.