



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

CR-B 2023 Cardiac Risk

INSTITUTION: LKF Laboratorium fur Klinische Forschung GmbH
Schwentinental GE 24223


ATTENTION: Volker El-Samalouti PhD

CAP NUMBER: 7234136-01 **Kit#** 1

KIT INFORMATION:	Kit ID:	Kit Mailed:	Original Evaluation:	Next Mailing Date:
	36423967	10/30/2023	12/1/2023	3/25/2024

COPIED TO: CAP

LEGEND: Exception Reason Codes appearing in this evaluation:
<NONE>




Reviewed By



Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 7234136-01

Kit# 1

Institution: LKF Laboratorium fur Klinische Forschung GmbH

Attention: Volker El-Samalouti PhD

City / State: Schwentimental GE 24223

Kit ID: 36423967

Kit Mailed: 10/30/2023

Original Evaluation: 12/1/2023

**EVALUATION
ORIGINAL**

CR-B 2023 Cardiac Risk

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics									Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+ 100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper		Your Grade	
hsCRP mg/L ROCHE	HSCR-04	5.8	5.60	0.33	476	+0.6	3.9	7.3	Acceptable	
	HSCR-05	0.2	0.29	0.09	411	-1.0	0.2	0.4	Acceptable	
	HSCR-06	2.5	2.62	0.23	477	-0.5	1.8	3.5	Acceptable	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.