



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

CR-A 2024 Cardiac Risk

INSTITUTION: LKF Laboratorium fur Klinische Forschung GmbH
Schwentinental GE 24223

ATTENTION: Volker El-Samalouti PhD

CAP NUMBER: 7234136-01 **Kit#** 1

KIT INFORMATION:	Kit ID: 37734350	Kit Mailed: 3/25/2024	Original Evaluation: 4/22/2024	Next Mailing Date: 10/28/2024
-------------------------	----------------------------	---------------------------------	--	---

COPIED TO: CAP

LEGEND: **Exception Reason Codes appearing in this evaluation:**
<NONE>

S. J. [Signature]

Reviewed By

1710612024

Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 7234136-01

Kit# 1

Institution: LKF Laboratorium fur Klinische Forschung GmbH

Attention: Volker El-Samalouti PhD

City / State: Schwentimental GE 24223

Kit ID: 37734350

Kit Mailed: 3/25/2024

Original Evaluation: 4/22/2024

**EVALUATION
ORIGINAL**

CR-A 2024 Cardiac Risk

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics									Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper		Your Grade	
hsCRP mg/L ROCHE	HSCR-01	3.5	3.70	0.30	480	-0.7	2.5	4.9	Acceptable	
	HSCR-02	13.6	13.74	0.52	469	-0.3	9.6	17.9	Acceptable	
	HSCR-03	1.4	1.44	0.18	476	-0.2	1.0	1.9	Acceptable	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.